



**INTERNATIONAL STANDARD ISO 7393-1:1985**  
TECHNICAL CORRIGENDUM 1

Published 2001-03-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

## **Water quality — Determination of free chlorine and total chlorine —**

### **Part 1: Titrimetric method using *N,N*-diethyl-1,4-phenylenediamine**

#### **TECHNICAL CORRIGENDUM 1**

*Qualité de l'eau — Dosage du chlore libre et du chlore total —*

*Partie 1: Méthode titrimétrique à la *N,N*-diéthylphénylène-1,4 diamine*

*RECTIFICATIF TECHNIQUE 1*

Technical Corrigendum 1 to International Standard ISO 7393-1:1985 was prepared by Technical Committee ISO/TC 147, *Water quality*, Subcommittee SC 2, *Physical, chemical and biochemical methods*.

*Page 2*

In **4.5.2 Standardization of the solution**, replace the existing formula with the following:

$$c_1 = V_2 \frac{c_2}{2V_1}$$

**Pour plus d'infos, merci de nous contacter.**

**Association Sénégalaise de Normalisation**

4 Avenue Jean Jaurès, Immeuble El Hadj Omar DIA, 6 ème Etage  
Dakar - SENEGAL

**Tel: +221 33 829 58 25**

**Email: [asn@asn.sn](mailto:asn@asn.sn)**