



INTERNATIONAL STANDARD ISO 10155:1995
TECHNICAL CORRIGENDUM 1

Published 2002-02-15

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • МЕЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

Stationary source emissions — Automated monitoring of mass concentrations of particles — Performance characteristics, test methods and specifications

TECHNICAL CORRIGENDUM 1

Émissions de sources fixes — Contrôle automatique des concentrations en masse de particules — Caractéristiques de fonctionnement, modes opératoires d'essai et spécifications

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to International Standard ISO 10155:1995 was prepared by Technical Committee ISO/TC 146, *Air quality*, Subcommittee SC 1, *Stationary source emissions*.

Page 9

In annex A, clause **A.3 Determination of the confidence interval for the linear regression**

Delete equation (A.11) and replace with: $S = \left(\frac{S_{yy}}{n - 2} \right)^{1/2} \left(1 - \frac{S_{xy}^2}{S_{xx} S_{yy}} \right)^{1/2}$

Pour plus d'infos, merci de nous contacter.

Association Sénégalaise de Normalisation

4 Avenue Jean Jaurès, Immeuble El Hadj Omar DIA, 6 ème Etage
Dakar - SENEGAL

Tel: +221 33 829 58 25

Email: asn@asn.sn